

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



1736

Accredited to  
ISO/IEC 17025:2017

### Pathology Management Services

Issue No: 029 Issue date: 17 June 2019

Rose Mill  
Union Street  
Middleton  
Manchester  
M24 6DD

Contact: Mrs B Andrew  
Tel: +44 (0)161 643 5330  
Fax: +44 (0)161 653 3647  
E-Mail: path.man@zen.co.uk  
Website: www.pathology-management.co.uk

Testing performed at the above address only

#### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
ENVIRONMENTAL SWABS	<p><u>Microbiological Tests</u></p> <p>Enumeration:</p> <p>Aerobic Colony Count</p> <p><i>Bacillus cereus</i> (presumptive)</p> <p>Enterobacteriaceae (presumptive)</p> <p><math>\beta</math>-glucuronidase positive <i>Escherichia coli</i></p> <p><i>Listeria</i> spp including confirmation of <i>L. monocytogenes</i> and other species identification if requested</p> <p><i>Pseudomonas</i> spp (presumptive)</p> <p>Coagulase positive Staphylococci including <i>Staphylococcus aureus</i></p> <p>Yeasts and Moulds</p>	<p>Documented In-House Methods</p> <p>SOP-MB-001 using a pour plate technique on PCA agar at 30°C for 48 hours</p> <p>SOP-MB-005 using colony count technique on PEMBA agar at 30 °C for 24hrs, and haemolysis confirmation using sheep blood agar.</p> <p>SOP-MB-003 using a colony count technique at 37 °C based on BS EN ISO 21528-2:2017</p> <p>SOP-MB-015 based on BS ISO 16649-2:2001 using TBX</p> <p>SOP-MB-017.2 based on BS EN ISO 11290-2:2017 with confirmation using API Listeria</p> <p>SOP-MB-007 based on ISO 13720:2010</p> <p>SOP-MB-006 using colony count technique at 37 °C based on BS EN ISO 6888-1:1999</p> <p>SOP-MB-008 using a colony count technique on RBC agar at 25 °C</p>



1736

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

### United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

### Pathology Management Services

Issue No: 029 Issue date: 17 June 2019

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
ENVIRONMENTAL SWABS (cont'd)	<u>Microbiological Tests (cont'd)</u>  Detection:  <i>Listeria</i> spp including confirmation of <i>L monocytogenes</i> and other species identification if requested  <i>Salmonella</i> spp	Documented In-House Methods (cont'd)  SOP-MB-017 based on BS EN ISO 11290-1:2017 with confirmation using API Listeria  SOP-MB-016 using an enrichment technique and confirmation with API 20E based on BS EN ISO 6579-1:2017
FOOD and FOOD PRODUCTS including raw ingredients	<u>Microbiological Tests</u>  Enumeration:  Aerobic Colony Count  <i>Bacillus cereus</i> (presumptive)  Coliforms (presumptive)  <i>Clostridium perfringens</i>  Enterobacteriaceae (presumptive)  <i>Enterococcus</i> spp	SOP-MB-001 using a pour plate technique on PCA agar at 30°C for 48 hours  SOP-MB-005 using colony count technique on PEMBA agar at 30 °C for 24hrs, and haemolysis confirmation using sheep blood agar.  SOP-MB-002 using colony count technique at 30 °C for milk and milk-based products and 37 °C for general foods based on BS ISO 4832:2006  SOP-MB-009 using an anaerobic colony count technique at 37 °C based on BS EN ISO 7937:2017  SOP-MB-003 using a colony count technique at 37 °C based on BS EN ISO 21528-2:2017  SOP-MB-010 using colony count technique on Slanetz and Bartley agar at 37 °C and confirmation with Aesculin agar.



1736

Accredited to  
ISO/IEC 17025:2017

### Schedule of Accreditation

issued by

### United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

### Pathology Management Services

Issue No: 029 Issue date: 17 June 2019

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
FOOD and FOOD PRODUCTS including raw ingredients (cont'd)	<p><u>Microbiological Tests (cont'd)</u></p> <p>Enumeration: (cont'd)</p> <p><math>\beta</math>-glucuronidase positive <i>Escherichia coli</i></p> <p><i>Listeria</i> spp including confirmation of <i>L. monocytogenes</i> and other species identification if requested</p> <p><i>Pseudomonas</i> spp (presumptive)</p> <p>Coagulase positive Staphylococci including <i>Staphylococcus aureus</i></p> <p>Total Aerobic Mesophilic spore count</p> <p>Yeasts and Moulds</p> <p>Detection: <i>Escherichia coli</i>, (presumptive)</p> <p><i>Listeria</i> spp including confirmation of <i>L. monocytogenes</i> and other species identification if requested</p> <p><i>Salmonella</i> spp</p>	<p>SOP-MB-015 based on BS ISO 16649-2:2001 using TBX</p> <p>SOP-MB-017.2 based on BS EN ISO 11290-2:2017 with confirmation using API Listeria</p> <p>SOP-MB-007 based on BS ISO 13720:2010</p> <p>SOP-MB-006 using colony count technique at 37°C based on BS EN ISO 6888-1:1999</p> <p>SOP-MB-012 using colony count technique at 30°C based on BS 4285: Section 3.3:1986 (withdrawn, not replaced)</p> <p>SOP-MB-008 using colony count technique on RBC agar at 25 °C</p> <p>SOP-MB-018 using a single tube enrichment technique based on BS EN ISO 7251:2005</p> <p>SOP-MB-0017 based on BS EN ISO 11290-1:2017 with confirmation using API Listeria</p> <p>SOP-MB-016 using an enrichment technique based and confirmation with API 20E based on BS EN ISO 6579-1:2017</p>



1736

Accredited to  
ISO/IEC 17025:2017

### Schedule of Accreditation

issued by

### United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

### Pathology Management Services

Issue No: 029 Issue date: 17 June 2019

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>WATERS</p> <p>Natural, potable, pools and spas (unless otherwise stated)</p>	<p><u>Microbiological Tests</u></p> <p>Isolation and Enumeration:</p> <p><i>Clostridium perfringens</i> and other sulphite reducing <i>clostridia</i></p> <p>Coliforms (presumptive)</p> <p><i>Escherichia coli</i></p>	<p>Documented In-House Methods based on "Microbiology of Drinking Water", Environment Agency (unless otherwise stated)</p> <p>SOP-MB-026 by membrane filtration based on MDW (2015) Part 6</p> <p>SOP-MB-029 by membrane filtration, single membrane based on MDW (2016) Part 4B</p> <p>SOP-MB-029 by membrane filtration, single membrane based on MDW (2016) Part 4B</p>
<p>Natural and potable only</p>	<p>Enterococci (Faecal streptococci)</p> <p><i>Pseudomonas aeruginosa</i></p> <p>Total viable aerobic organisms</p>	<p>SOP-MB-027 by membrane filtration based on MDW (2012) Part 5A</p> <p>SOP-MB-025 by membrane filtration based on MDW (2015) Part 8B</p> <p>SOP-MB-022 using a colony count technique at 22 °C, 30 °C and/or 37 °C based on MDW (2012) Part 7A</p>
<p>Pools and spas only</p>	<p><i>Staphylococcus aureus</i></p>	<p>SOP-MB-028 by membrane filtration onto Mannitol Salt agar and confirmation by Coagulase test and DNase</p>
<p>WATERS – Natural, Potable, Cooling Towers, Spas, Showers</p>	<p>Detection, enumeration and confirmation:</p> <p><i>Legionella pneumophila</i> Serogroups 1, 2-14, and/or presumptive <i>Legionella</i> spp (not <i>L pneumophila</i>)</p>	<p>SOP-MB-021 based on BS EN ISO 11731:2017 using filtration of 500ml volumes with washing and treatment, plated on GVPC and identification by latex agglutination (Matrix A &amp; B: Procedure 8, 9 &amp; 10, Media C)</p>

END