


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 Accredited to ISO/IEC 17025:2017	<b>Pathology Management Services</b>	
	Issue No: 037	Issue date: 05 July 2024
	Rose Mill Union Street Middleton Manchester M24 6DD	Contact: Mr D Andrew Tel: +44 (0)161 643 5330 Fax: +44 (0)161 653 3647 E-Mail: path.man@hotmail.com Website: www.pathology-management.co.uk
Testing performed at the above address only		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
ENVIRONMENTAL SWABS	<u>Microbiological Tests</u>  Enumeration:  Aerobic Colony Count  <i>Bacillus cereus</i> (presumptive)  Enterobacteriaceae (presumptive)  $\beta$ -glucuronidase positive <i>Escherichia coli</i>  <i>Listeria</i> spp including confirmation of <i>L. monocytogenes</i>  <i>Pseudomonas</i> spp (presumptive)  Coagulase positive Staphylococci including <i>Staphylococcus aureus</i>  Yeasts and Moulds	Documented In-House Methods  SOP-MB-001 using a pour plate technique on PCA agar at 30°C for 48h  SOP-MB-005 using colony count technique on PEMBA agar at 30 °C for 24h, and haemolysis confirmation using sheep blood agar  SOP-MB-003 using a colony count technique at 37 °C based on BS EN ISO 21528-2:2017  SOP-MB-015 based on BS ISO 16649-2:2001 using TBX  SOP-MB-017.2 based on BS EN ISO 11290-2:2017 with confirmation using API Listeria  SOP-MB-007 based on ISO 13720:2010  SOP-MB-006 using colony count technique at 37 °C based on BS EN ISO 6888-1:2021+A1:2023 with confirmation using Staphylase latex agglutination kit  SOP-MB-008 using a colony count technique on RBC agar at 25 °C



1736  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Pathology Management Services**  
**Issue No: 037    Issue date: 05 July 2024**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
ENVIRONMENTAL SWABS (cont'd)	<u>Microbiological Tests</u> (cont'd)  Detection:  <i>Listeria</i> spp including confirmation of <i>L. monocytogenes</i>  <i>Salmonella</i> spp	Documented In-House Methods (cont'd)  SOP-MB-017 based on BS EN ISO 11290-1:2017 with confirmation using API <i>Listeria</i>  SOP-MB-016 using an enrichment technique and confirmation with API 20E based on BS EN ISO 6579-1:2017+A1:2020
FOOD and FOOD PRODUCTS including raw ingredients	<u>Microbiological Tests</u>  Enumeration:  Aerobic Colony Count  <i>Bacillus cereus</i> (presumptive)  Coliforms (presumptive)  <i>Clostridium perfringens</i>  Enterobacteriaceae (presumptive)  <i>Enterococcus</i> spp	SOP-MB-001 using a pour plate technique on PCA agar at 30°C for 48h  SOP-MB-005 using colony count technique on PEMBA agar at 30 °C for 24h, and haemolysis confirmation using sheep blood agar.  SOP-MB-002 using colony count technique at 30 °C for milk and milk-based products and 37 °C for general foods based on BS ISO 4832:2006  SOP-MB-009 using an anaerobic colony count technique at 37 °C based on BS EN ISO 15213-2:2023 with confirmation by acid phosphatase  SOP-MB-003 using a colony count technique at 37 °C based on BS EN ISO 21528-2:2017  SOP-MB-010 using colony count technique on Slanetz and Bartley agar at 37 °C and confirmation with Aesculin agar



1736  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Pathology Management Services**  
**Issue No: 037    Issue date: 05 July 2024**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
FOOD and FOOD PRODUCTS including raw ingredients (cont'd)	<u>Microbiological Tests</u> (cont'd)  Enumeration: (cont'd)  $\beta$ -glucuronidase positive <i>Escherichia coli</i>  <i>Listeria</i> spp including confirmation of <i>L. monocytogenes</i>  <i>Pseudomonas</i> spp (presumptive)  Coagulase positive Staphylococci including <i>Staphylococcus aureus</i>  Yeasts and Moulds  Detection:  <i>Escherichia coli</i> , (presumptive)  <i>Listeria</i> spp including confirmation of <i>L. monocytogenes</i>  <i>Salmonella</i> spp	Documented In-House Methods (cont'd)  SOP-MB-015 based on BS ISO 16649-2:2001 using TBX  SOP-MB-017.2 based on BS EN ISO 11290-2:2017 with confirmation using API Listeria  SOP-MB-007 based on BS EN ISO 13720:2010  SOP-MB-006 using colony count technique at 37°C based on BS EN ISO 6888-1:2021+A1:2023 with confirmation using Staphylase latex agglutination kit  SOP-MB-008 using colony count technique on RBC agar at 25°C  SOP-MB-018 using a single tube enrichment technique based on BS EN ISO 7251:2005  SOP-MB-0017 based on BS EN ISO 11290-1:2017 with confirmation using API Listeria  SOP-MB-016 using an enrichment technique based and confirmation with API 20E based on BS EN ISO 6579-1:2017 +A1:2020



1736  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Pathology Management Services**  
**Issue No: 037 Issue date: 05 July 2024**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
WATERS Natural, potable, pools and spas (unless otherwise stated)	<u>Microbiological Tests</u> (cont'd)  Enumeration:  <i>Clostridium perfringens</i> and other sulphite reducing <i>clostridia</i>  Coliforms (presumptive)  <i>Escherichia coli</i>	Documented In-House Methods based on "Microbiology of Drinking Water", Environment Agency (unless otherwise stated)  SOP-MB-026 by membrane filtration based on MDW (2021) Part 6 with confirmation by acid phosphatase  SOP-MB-029 by membrane filtration, single membrane based on MDW (2016) Part 4B  SOP-MB-029 by membrane filtration, single membrane based on MDW (2016) Part 4B
Natural and potable only	Enterococci (Faecal streptococci)  <i>Pseudomonas aeruginosa</i>	SOP-MB-027 by membrane filtration based on MDW (2012) Part 5A  SOP-MB-025 by membrane filtration based on MDW (2020) Part 8B
Pools and spas only	Total viable aerobic organisms	SOP-MB-022 using a colony count technique at 22 °C, and/or 37 °C based on MDW (2012) Part 7A
	Coagulase positive Staphylococci including <i>Staphylococcus aureus</i>	SOP-MB-028 in-house method by membrane filtration onto Mannitol Salt agar and confirmation using Staphylase latex agglutination kit
WATERS – Natural, Potable, Cooling Towers, Spas, Showers	<i>Legionella pneumophila</i> Serogroups 1, 2-14, and/or presumptive <i>Legionella</i> spp (not <i>L. pneumophila</i> )	SOP-MB-021 based on BS EN ISO 11731:2017 using filtration of 500ml volumes with washing and treatment, plated on GVPC and identification by latex agglutination (Matrix A & B: Procedure 8, 9 & 10, Media C)
END		